



Ceveron TGA RB Kit

LOT B336C00

 2024-07-31

3013831- 017 20/11/2023

Ceveron 100 series barcode



Kontrolle control	Lot	Parameter	Einheit unit	Richtwert (Vertrauensbereich) expected value (confidence range)	Std.
Ceveron TGA Control High	1Y33C00	Peak thrombin	nM	195 (150 - 240)	IS 19/188
Ceveron TGA Control Low	2734C00			120 (92 - 148)	IS 19/188